

Volumetric flasks, class A, white graduation, works certified



PYREX®, borosilicate glass, with PE stopper

- Calibrated to the tolerances required by the United States Pharmacopoeia (USP), which are tighter than those of ISO/DIN standards
- Each flask is laser etched with individual serial number and date of manufacture
- Calibrated to contain (TC, In)
- White enamel inscriptions

Complies with ISO 1042 and DIN 12664.

Ordering information: Supplied with certificate of calibration traceable to National Standards.

Capacity (ml)	NS	Tolerance (± ml)	Pk	Cat. No.
1	7/16	0,02	5	612-3200
2	7/16	0,02	5	612-3201
5	10/19	0,02	5	612-3232
10	10/19	0,02	5	612-3239
20	10/19	0,03	5	612-3202
25	10/19	0,03	5	612-3237
50	12/21	0,05	5	612-3234
100	14/23	0,08	5	612-3220
200	14/23	0,1	5	612-3214
250	14/23	0,12	2	612-3199
500	19/26	0,15	2	612-3229
1000	24/29	0,3	2	612-3227
2000	29/32	0,5	1	612-3225
5000	34/35	1	1	612-3223

Volumetric flasks, class A, white graduation, BLAUBRAND®

Brand



BLAUBRAND®, conformity certified, DURAN®, with PP stopper

- Calibrated to contain (TC, In)
- Marks and inscriptions in high contrast white enamel
- Conformity certified

DIN EN ISO 1042

Ordering information: Supplied with a batch certificate. Volumetric flasks also available on request with an individual certificate, a USP individual certificate or a DAkkS calibration certificate.

Capacity (ml)	NS	Tolerance (± ml)	Wide mouth	Pk	Cat. No.
5	10/19	0,04	√	2	612-5084
10	10/19	0,04	√	2	612-5085
20	10/19	0,04	-	2	612-5086
25	10/19	0,04	-	2	612-5087
50	12/21	0,06	-	2	612-5088
50	14/23	0,10	√	2	612-5089
100	14/23	0,10	-	2	612-5090
200	14/23	0,15	-	2	612-5091
250	14/23	0,15	-	2	612-5092
500	19/26	0,25	-	2	612-5093
1000	24/29	0,40	-	1	612-5094



VWR COLLECTION

Delivering high quality and superior performance

Please contact your local VWR sales office or go to vwr.com for the latest news and special offers